

1. Title

Evaluation of the Point of Care Molecular Diagnostic Genedrive HCV ID Kit for the detection of HCV RNA in clinical samples.

Author names and affiliations.

Abhishek Padhi^a

Department of Clinical Virology, ILBS, New Delhi, India. 110070^a

e-mail : abhi038450@gmail.com

Ekta Gupta^a

Department of Clinical Virology, ILBS, New Delhi, India. 110070^a

e-mail : ektagaurisha@gmail.com

Gaurav Singh^a

Department of Clinical Virology, ILBS, New Delhi, India. 110070^a

e-mail : gaurav_singh985@yahoo.com

Reshu Agarwal^a

Department of Clinical Virology, ILBS, New Delhi, India. 110070^a

e-mail : agarwalreshu286@gmail.com

Manoj Kumar Sharma^b

Department of Hepatology, ILBS, New Delhi, India. 110070^b

e-mail : manojkumardm@gmail.com

This peer-reviewed article has been accepted for publication but not yet copyedited or typeset, and so may be subject to change during the production process. The article is considered published and may be cited using its DOI 10.1017/S0950268820002812.

Epidemiology & Infection is published by Cambridge University Press.

Shiv Kumar Sarin^b

Department of Hepatology, ILBS, New Delhi, India. 110070^b

e-mail : shivsarin@gmail.com

Corresponding author

Ekta Gupta^a

Department of Clinical Virology, ILBS, New Delhi, India. 110070^a

e-mail : ektagaurisha@gmail.com

Accepted Manuscript

2. Abstract

Despite the availability of an effective treatment for HCV infection, major bottle neck in the HCV care cascade is difficulty in getting the HCV RNA done on serologically positive (antibody to HCV) patients. This is due to the limited availability of HCV RNA test in the periphery. Nucleic acid amplification testing at point of care might revolutionise the HCV care continuum through its increased sensitivity, decreased turn around time and ease of performance in the field.

Genedrive is one such assay which has recently got a CE- IVD certification. The diagnostic accuracy of Genedrive HCV ID kit for the qualitative detection of HCV RNA was evaluated by comparing with Abbott HCV RNA in an Indian demographic setting and across a range of different genotypes commonly found in India. For the assay evaluation 150 HCV RNA positive and 170 HCV RNA negative samples as tested by Abbott HCV RNA assay (Abbott, Wiesbaden, Germany) on Automated m2000sp/m2000rt platform, as per the manufacturer's instructions, that were retrieved from -80°C in the Virology repository were included in the study. All the samples were retrieved and re tested for HCV RNA on Genedrive HCV ID kit and compared with Abbott HCV RNA results.

Comparison of the Genedrive HCV ID kit with the Abbott Real Time HCV assay revealed a sensitivity of 100 % (95% CI 97.9 to 100) and a specificity of 100 % (95% CI 97.9 to 100) with a positive predictive value and negative predictive value of 100%. Overall diagnostic accuracy of Genedrive was found to be 100% (95%

CI 98.9 to 100). This study demonstrates that Genedrive HCV ID kit can be used for decentralised testing of HCV and ultimately finding the missing million.

Keywords

Hepatitis C, Point of care test, HCV care cascade

3. Background

Hepatitis C virus (HCV) infection known for its chronicity resulted in approximately 3,99 000 death in 2016 due to liver cirrhosis and hepatocellular carcinoma [1]. Globally an estimated 71 million individuals are infected with HCV out of which India is one of the six countries (others being China, Pakistan, Nigeria, Egypt & Russia) where more than 50% of the infected population resides[2] . A recent meta-analysis of HCV infection in Indian population predicted around 3 – 9 million persons with active Hepatitis C infection[3].

In 2016, the world health organization (WHO) launched its ambitious strategy to eliminate viral hepatitis as a public threat by 2030 with a targets of 90% reduction in incident cases of hepatitis B and C and a 65% reduction in mortality[4]. In absence of vaccine against HCV and with the introduction of directly acting antivirals (DAA) as an effective treatment option, the onus squarely lies on an accurate diagnostic modality to achieve this target.

The traditional approach to HCV diagnosis requires an HCV antibody test followed by detection of HCV RNA to confirm viremia in HCV antibody (HCV-ab) positive cases[1]. However, HCV-Ab detection includes individuals who have cleared infection either spontaneously or through treatment (estimating exposure as well as active infection), and also includes false positives[5]. Furthermore low signal upon cut off presents a diagnostic dilemma in the detection of HCV-ab[6]. Again the rapid diagnostic tests pitted as the point of care test (POCT) have varied sensitivity in different settings [7][8]. As a result, there has been a paradigm shift towards using HCV PCR to determine accurately the population prevalence of active infection[5].

While HCV RNA testing allows determining current status of infection thus indicating individuals for treatment[9], it is confined to few high-end laboratories. This centralized arrangement of HCV RNA testing may result in patients losing to the follow up. Presently, the Cepheid Xpert HCV Viral Load assay (Cepheid) is one of the CE-IVD certified assay for decentralized HCV VL determination. Though the Xpert HCV reported to have a good performance, a high sample volume and need of electrical power supply limits its field use[10]. Hence a robust, fully automated, high throughput platform with a low turn around time is the need of the hour that can serve as a single point molecular testing and refer to care centre so that loss to follow up can be prevented particularly in the grass root level.

The Genedrive HCV ID Kit is one such platform which is European Union CE approved, hand held battery operated fully automated platform developed as a point-of-need molecular test for confirmation of chronic HCV. The test is a cartridge based reverse transcriptase polymerase chain reaction (RT-PCR) with an end point melt analysis of fluorescent molecular probes. It has so far been validated in European and African settings with the predominant HCV genotype as genotype 1,3,4 and 5 , and has demonstrated good diagnostic accuracy (sensitivity 99.8 %, specificity 100%) [10]. The Genedrive HCV ID Kit requires validation in Indian context where the circulation of HCV genotype is predominately 3 and 1 to ensure suitability for the Indian population.

4. Objectives

The objectives of this study were to determine the diagnostic accuracy of the Genedrive HCV ID Kit for HCV RNA detection, as a confirmatory test for seropositive HCV patients in Indian demographic settings.

5. Study Design

5.1. Samples

Samples were retrieved from -80°C from the Virology data base, 150 positive samples for HCV RNA and 170 negative samples for HCV RNA . All reference test data were collected retrospectively, whilst Genedrive data were collected prospectively. All specimens were remnants from routine testing of participants either coming to the outpatient department or admitted in the hospital. The specimens were plasma derived from blood collected in EDTA vacutainers that had been stored at -80°C and were collected at ILBS between date (October 2017) and date (Feb 2019).

Cases were determined as serologically positive for HCV with confirmation of HCV RNA presence by the Abbott Real Time HCV assay (Reference Test, henceforth called Abbot assay). Whilst the controls were determined as serologically positive for HCV with confirmation of HCV RNA absence by the Abbott assay (Reference Test).

HBV and HIV co-infected samples and samples with inadequate volume were excluded from the analysis.

The study was performed blind folded with samples run parallelly in both Abbott Real Time HCV assay and Genedrive HCV ID Kit. The supervisor of the experiment knew the HCV RNA results of Abbott. At the time of experiment the technical staff who performed both the assays in parallel (Genedrive and repeat Abbott) was unaware of the results.

5.2. Abbott Real Time HCV assay

The Abbott assay was considered as reference test for detection of HCV RNA in plasma samples. The test is quantitative with a 95% limit of detection (LOD) of 12 IU/mL and measuring range of 12 to 10^8 IU/mL. 750 μ L of plasma per specimen was analysed using the Abbott assay on the automated m2000sp/m2000rt platform, as per the manufacturer's instructions.

5.3. Genedrive HCV ID Kit

The Genedrive assay has a lower limit of Detection of $3.37 \log_{10}$ IU/mL (2362 IU/mL) and kit has been validated to work on all the genotypes (as per the kit literature). [11].

5.4. HCV Genotyping

Wherever a viral load of $>3 \log_{10}$ value was their the genotyping was done.

Genotyping of positive HCV RNA samples was performed as described previously [12]. Briefly, genotyping of HCV RNA positive samples was performed using the 5'UTR region of the viral RNA. A further confirmation of genotype 1 samples was performed using the NS5B RNA region [13]. Genotyping of amplified and purified DNA was performed by bidirectional sequencing using ABI Big Dye chemistry on the ABI 3500Dx series genetic analyzer (Life Technologies, Waltham, MA).

Sequence reads were aligned using DNA Baser v3.5.1 (Heracle BioSoft SRL,

Romania). A Basic Local Alignment Search Tool (BLAST) was performed using obtained sequences using the database of NCBI.

5.5. Genedrive assay workflow

The Genedrive HCV assay is a two step procedure initially requiring plasma preparation followed by a reverse transcription polymerase chain reaction. In the Genedrive assay 30 µl of plasma sample was initially diluted in 60 µl of nuclease free water (1:2 ratio). Following which 15µL of the diluted plasma was added to each of the three channels of the plasma preparation cartridge. The protease incubation step was then performed within the Genedrive instrument, consisting of a 5 min incubation at 37°C followed by 5 min at 95°C, thus rendering the sample non-infectious. During this incubation time, 100µL of nuclease free water was added to resuspend the RT-PCR reagents. After removal of the cartridge from the device, 30 µL of the RT-PCR suspension was dispensed into each channel, which was resealed with a new lid. The cartridge was placed back in the Genedrive instrument, and the programme resumed. The target RNA was converted to DNA (via RT) and subsequently amplified by asymmetric PCR. Following amplification, a fluorescent probe was used to detect target-specific sequences by monitoring changes in fluorescent signal intensity that occurred during dissociation of the fluorescent probe from its hybridised target sequence (if present) as temperature increased. The Genedrive instrument detects this fluorescence at

defined melt temperature positions (T_m) for both the HCV target and for the IPC.

The instrument has a run time of 88 minute.

Genedrive is an user friendly device and requires minimal technical expertise to operate.

The results are interpreted automatically by the Genedrive instrument as Detected, Positive and Undetected, Negative. At times Indeterminate result is displayed by the Genedrive assay. As per the kit insert an indeterminate result implies the detection of HCV RNA only in one of the channels. In such conditions the same plasma sample should be retested. If this is not possible, or if a second indeterminate result is obtained, a fresh plasma sample from the same patient should be obtained.

5.6. Statistical analysis

Statistical analysis was done using the Statistical SPSS software, version 22.0; Chicago, IL, USA). Continuous variables were expressed as mean \pm SD or median (range) as appropriate and categorical variables were expressed as percentage. All statistical tests were 2-tailed, and results were statistically significant with a p value was less than 0.05.

6. Results

6.1. Sample details

A total of 320 samples were included in the study, the baseline characteristics of the study population is described. Of the 150 HCV positive specimens, thirteen were below the Genedrive 95 % Limit of Detection of $3.37 \log_{10}$ IU/mL (2362 IU/mL), whilst the remaining 137 were above. Genotyping could be done in 105 samples.

Table 1 – Baseline characteristics of the study population

Variable	Number
Age (mean \pm SD) years	47.1 \pm 13.5
Male : female	1.8:1
Genotype (n=105)	
1	35 (33.3%)
3	68 (64.7%)
4	2 (1.9%)
Median RNA levels (\log_{10} IU/mL) (Range)	5.66 (2.01 – 7.47)

6.2. Diagnostic accuracy

Comparison of the Genedrive HCV ID Kit with the Abbott Real Time HCV assay revealed a sensitivity of 100% (95% CI 97.6 to 100.0) and a specificity of 100%

(95 % CI 97.9 to 100), with a Positive Predictive Value and an Negative Predictive Value of 100%. The overall diagnostic accuracy of the Genedrive was determined at 100% (95% CI 98.9 to 100.0). Thirteen specimens were included in the study below the stated Genedrive 95% Limit of Detection of 2362 IU/ml (3.37 log₁₀ IU/mL), these specimens had a median viral load of 2.96 log₁₀ IU/mL and a range of 2.01 – 3.32 log₁₀ IU/mL. All thirteen specimens were detected by Genedrive. The lowest viral load that was detected by genedrive was 103 IU/ml (2.01 log₁₀IU/ml).

Table 2 – Comparison of HCV RNA detection between Abbott and Genedrive assay.

Abbott			
Undetected	<3.37 log ₁₀	>3.37 log ₁₀	Total

		IU/mL		IU/mL	
	Undetectd	170	0	0	170
Genedrive	Detected	0	13	137	150
	Total	170	13	137	320

6.3. Discrepant samples

No discrepancy was observed between Genedrive with the Abbott HCV assay.

7. Discussion

The present study aimed at evaluating the diagnostic accuracy of Genedrive HCV assay for the qualitative detection of HCV RNA. A comparison of Genedrive was done with Abbott Real Time HCV test, which offers viral load and genotype testing and is currently considered as the gold standard. Sensitivity of 100% (95% CI 97.6 to 100.0) and a specificity of 100% (95% CI 97.9 to 100) was observed in our study. Studies done by Libre *et al* [10] found high diagnostic sensitivity of 98.6% (95% CI 96.9% to 99.5%) and 100 % specificity (95% CI 99.3% to 100%) to detect HCV viraemia in Western and South African laboratories.

WHO recommends that HCV serology testing be offered to individuals who are part of a population with high HCV prevalence or who have a history of HCV risk exposure/ behaviour and confirm the diagnosis of chronic HCV infection by nucleic acid testing (NAT) for the detection of HCV RNA [9]. Currently HCV RNA testing is available in only few of a centralised medical set ups resulting in < 1% of infected individual being aware of their infection, especially in the low and middle income countries [14]. In Europe only 4.1% of individuals have been cured from initial viraemia [15] and the situation is even more grim in resource limited countries where it is <1 % [16][17]. In countries like Egypt and Mongolia the prevalence of HCV is very high but the treatment rate were appalling, 0.12% and 1.2% respectively, thereby suggesting that there was a negative correlation between HCV prevalence and treatment [18][19]. A national population- based screening program was initiated in Egypt in October 2018 to augment the treatment flow and meet the disease elimination targets for HCV infection. Availability of generic Directly acting antivirals (DAAs) has been a game changer in the treatment of HCV in Egypt. Nearly 50 million people participated in screening, and approximately 1 million patients were treated for HCV infection [20] [21] [22] [23].

There is an urgency in the need of a point of care molecular test which can be an alternative to the conventional HCV NAATs, that can be used in the field thereby decentralising the HCV diagnosis mainly in the resource limited countries.

The Genedrive is a portable, fully integrated, PCR-based platform. Genedrive has been designed to enable 'Direct to PCR' analysis of sample matrices without the requirement to undertake pre-extraction of DNA, all in approximately 90 minutes from sample collection. Genedrive is a 600 gram handheld thermocycler that qualitatively detects HCV RNA by NAAT in plasma and serum samples. It fulfils the target product profile of HCV diagnosis drafted by Foundation for innovative new diagnostics (FIND)/WHO (specificity > 98% and sensitivity > 95%) and has shown diagnostic accuracy for all 6 major genotypes. As mentioned in the kit literature, the pan genotypic limit of detection (LoD) of Genedrive is 2362 IU/ml (95% CI 1966- 2758 IU/ml) as compared to Abbott HCV assay which is 12 IU/ml.

In our study we found few samples below the LoD being detected by the kit. This shows that there is a need to do large prospective study to actually verify the LOD of the genedrive assay. The present study analysed samples that were of Genotype 3, 4 and 1 as the most prevalent genotypes in India are 3 and 1 and most of the samples available with us in our archive belong to genotype 3 and 1. One of the limitations of this study is that genotypes 2, 5 and 6 were not tested.

Recent WHO guidelines suggest that "a limit of detection of 3000 IU/mL or lower would be acceptable and would identify 95% of those with viraemic infection"[24]. WHO recommends offering treatment to all individuals diagnosed with HCV infection who are 12 years of age or older, irrespective of disease stage

[9]. Furthermore with the advent of pan genotypic directly acting antivirals, need for genotyping prior to treatment holds no significance. Hence, Genedrive HCV ID Kit can potentially contribute to a decentralisation of clinical management of chronic hepatitis C, which may result in expansion of the treatment programme to rural areas of resource limited settings due to its relative ease of use in comparison to current centralised laboratory based systems for detection of HCV RNA.

Currently there are several NAAT point of care testings in pipeline and the Cepheid Xpert HCV viral Load assay is currently one of the CE- IVD certified assay for decentralised HCV viral load determination. Despite having good performance Xpert HCV test presents important limitations. The Xpert HCV assay requires large volume of sample (1 ml), a constant power supply, for which a basic laboratory set up is required. Moreover, the cartridge of Xpert HCV assay requires guanidium thiocyanate as lysis reagent which is highly toxic and requires special care while handling and disposal [12] [25]. Furthermore, the Xpert HCV Fingerstick assay is also available for the diagnosis of HCV infection and has the advantage of using whole blood as the sample. But the Xpert HCV Fingerstick assay also has the same limitations as that of the Xpert HCV assay. In contrast Genedrive is a battery operated handheld machine requiring minimal amount of sample (30µl) and there is no toxic chemicals which makes Genedrive ideal to be used in field settings.

The present study was carried out in a tertiary care centre and the test assay was performed by highly trained technicians in a controlled environment. Furthermore performance of Genedrive HCV assay requires a prospective validation in a real life decentralised peripheral resource limited setting in low- and middle income countries.

Finally, an important factor when assessing the feasibility of Point of care technologies is their cost, both for platform and assays. However, direct comparisons of costs to other platforms are challenging as there are geographic differences and unknowns like subsidised or negotiated pricing, distributor margins and duty . Furthermore, additional costs associated with the platform should be considered, including installation, maintenance, training and lifespan of the equipment and tests. While the ultimate Genedrive costs remain to be defined, the instrument is approximately \$5000, which is considerably less than other molecular systems, and every Genedrive HCV test are going to be available for \$30–\$40, counting on the country or region in question.

8. Conclusions

Despite the lack of quantification of HCV viral load Genedrive HCV assay offers a great potential to be used as a point of care NAAT thereby enabling real time

diagnosis, minimising loss to follow up and prompt management of patients with chronic HCV infection in any clinical setting.

Data Availability Statement

Readers may contact the authors for accessing data used in this study.

CRedit authorship contribution statement

Abhishek Padhi : Writing - original draft, Writing - review & editing, Methodology. **Ekta Gupta** - Conceptualization, Formal analysis, Investigation, Data curation, Writing - original draft, Writing - review & editing, Supervision, Project administration. **Gaurav Singh** : Resources, Methodology. **Reshu Agarwal** : Data curation. **M Sharma** : Writing - review & editing, Supervision. **Shiv Kumar Sarin** : Project administration, Supervision, Writing - review & editing.

Competing Interest

None declared

Funding

None

Acknowledgement

We thank Genedrive plc for supply of the instrument and Genedrive® HCV ID Kit.

Ethics

Since no personal data were collected and left over clinical samples sent to virology laboratory for routine diagnosis for HCV was used, ethical approval and individual consent were not applicable.

9. References

- [1] Hepatitis C, (n.d.). <https://www.who.int/news-room/fact-sheets/detail/hepatitis-c> (accessed February 14, 2020).
- [2] Polaris Observatory HCV Collaborators, Global prevalence and genotype distribution of hepatitis C virus infection in 2015: a modelling study, *Lancet Gastroenterology and Hepatology*. 2 (2017) 161–176. [https://doi.org/10.1016/S2468-1253\(16\)30181-9](https://doi.org/10.1016/S2468-1253(16)30181-9).
- [3] A. Goel, N. Seguy, R. Aggarwal, Burden of hepatitis C virus infection in India: A systematic review and meta-analysis, *Journal of Gastroenterology and Hepatology*. 34 (2019) 321–329. <https://doi.org/10.1111/jgh.14466>.

- [4] WHO-HIV-2016.06-eng.pdf [Internet]. [cited 2020 Feb 12]. Available from: <https://apps.who.int/iris/bitstream/handle/10665/246177/WHO-HIV-2016.06-eng.pdf> - Google Search, (n.d.). <https://www.google.com/search?q=WHO-HIV-2016.06-eng.pdf+%5BInternet%5D.+%5Bcited+2020+Feb+12%5D.+Available+from%3A+https%3A%2F%2Fapps.who.int%2Firis%2Fbitstream%2Fhandle%2F10665%2F246177%2FWHO-HIV-2016.06-eng.pdf&oq=WHO-HIV-2016.06-eng.pdf+%5BInternet%5D.+%5Bcited+2020+Feb+12%5D.+Available+from%3A+https%3A%2F%2Fapps.who.int%2Firis%2Fbitstream%2Fhandle%2F10665%2F246177%2FWHO-HIV-2016.06-eng.pdf&aqs=chrome..69i57.1603j0j7&sourceid=chrome&ie=UTF-8> (accessed February 14, 2020).
- [5] E. Adland, G. Jesuthasan, L. Downs, V. Wharton, G. Wilde, A.L. McNaughton, J. Collier, E. Barnes, P. Klenerman, M. Andersson, K. Jeffery, P.C. Matthews, Hepatitis virus (HCV) diagnosis and access to treatment in a UK cohort, *BMC Infectious Diseases*. 18 (2018) 461. <https://doi.org/10.1186/s12879-018-3367-3>.
- [6] N. Ballani, E. Gupta, Low signal-to-cutoff ratio (S/Co) in the diagnosis of hepatitis C: A diagnostic dilemma?, *Indian Journal of Gastroenterol*. 34 (2015) 413–414. <https://doi.org/10.1007/s12664-015-0593-0>.
- [7] A. Mane, J. Sacks, S. Sharma, H. Singh, A. Tejada-Strop, S. Kamili, K. Kacholia, R. Gautam, M. Thakar, R.S. Gupta, R. Gangakhedkar, Evaluation of five rapid diagnostic tests for detection of antibodies to hepatitis C virus (HCV): A step towards scale-up of HCV screening efforts in India, *PLoS ONE*. 14 (2019) e0210556. <https://doi.org/10.1371/journal.pone.0210556>.
- [8] S. Chevaliez, J.-M. Pawlotsky, New virological tools for screening, diagnosis and monitoring of hepatitis B and C in resource-limited settings, *Journal of Hepatology*. 69 (2018) 916–926. <https://doi.org/10.1016/j.jhep.2018.05.017>.
- [9] WHO | Guidelines for the screening, care and treatment of persons with hepatitis C infection, WHO. (n.d.). <http://www.who.int/hepatitis/publications/hepatitis-c-guidelines/en/> (accessed October 18, 2019).
- [10] A. Llibre, Y. Shimakawa, E. Mottez, S. Ainsworth, T.-P. Buivan, R. Firth, E. Harrison, A.R. Rosenberg, J.-F. Meritet, A. Fontanet, P. Castan, A. Madejón, M. Laverick, A. Glass, R. Viana, S. Pol, C.P. McClure, W.L. Irving, G. Miele, M.L. Albert, D. Duffy, Development and clinical validation of the Genedrive point-of-care test for qualitative detection of hepatitis C virus, *Gut*. 67 (2018) 2017–2024. <https://doi.org/10.1136/gutjnl-2017-315783>.
- [11] HCV-TPP-Report_17July2015_final.pdf, (n.d.). https://www.finddx.org/wp-content/uploads/2016/01/HCV-TPP-Report_17July2015_final.pdf (accessed October 18, 2019).
- [12] E. Gupta, P. Agarwala, G. Kumar, R. Maiwall, S.K. Sarin, Point -of -care testing (POCT) in molecular diagnostics: Performance evaluation of GeneXpert

- HCV RNA test in diagnosing and monitoring of HCV infection, *Journal of Clinical Virology*. 88 (2017) 46–51. <https://doi.org/10.1016/j.jcv.2017.01.006>.
- [13] D.G. Murphy, B. Willems, M. Deschênes, N. Hilzenrat, R. Mousseau, S. Sabbah, Use of sequence analysis of the NS5B region for routine genotyping of hepatitis C virus with reference to C/E1 and 5' untranslated region sequences, *Journal of Clinical Microbiology*. 45 (2007) 1102–1112. <https://doi.org/10.1128/JCM.02366-06>.
- [14] HepatitisC_Strategy_WEB.pdf, (n.d.). https://www.finddx.org/wp-content/uploads/2016/02/HepatitisC_Strategy_WEB.pdf (accessed October 28, 2019).
- [15] H. Razavi, S. Robbins, S. Zeuzem, F. Negro, M. Buti, A.-S. Duberg, F. Roudot-Thoraval, A. Craxi, M. Manns, R.T. Marinho, B. Hunyady, M. Colombo, S. Aleman, K. Antonov, P. Arkkila, K. Athanasakis, S. Blach, M. Blachier, A.J. Blasco, F. Calinas, J.L. Calleja, P.B. Christensen, M.E. Cramp, E. Croes, R.J. de Knegt, V. de Ledinghen, J.-M. Delile, C. Estes, K. Falconer, M. Färkkilä, R. Flisiak, S. Frankova, I. Gamkrelidze, J. García-Samaniego, J. Genov, J. Gerstoft, L. Gheorghe, A. Goldis, I. Gountas, S. Gregorčič, M. Gschwantler, J. Gunter, W. Halota, L. Harcouet, C. Hézode, P. Hoffmann, G. Horvath, I. Hrstic, P. Jarčuška, D. Jelev, A. Jeruma, M. Kåberg, J. Kieran, L.A. Kondili, I. Kotzev, H. Krarup, P. Kristian, M. Lagging, W. Laleman, P. Lázaro, V. Liakina, B. Lukšić, M. Maimets, M. Makara, L. Mateva, M. Maticic, F.S. Mennini, R. Mitova, C. Moreno, J. Mossong, K. Murphy, H. Nde, V. Nemecek, D. Nonkovic, S. Norris, M. Oltman, A.L.H. Øvrehus, G. Papatheodoridis, K. Pasini, D. Razavi-Shearer, K. Razavi-Shearer, H.W. Reesink, T. Reic, B. Rozentale, S.D. Ryder, R. Salupere, R. Sarmiento-Castro, C. Sarrazin, J.D. Schmelzer, I. Schréter, C. Seguin-Devaux, K. Simojoki, M. Simonova, P.J. Smit, K. Souliotis, D. Speiciene, J. Sperl, P. Stärkel, D. Struck, V. Sypsa, L. Thornton, I. Tolmane, K. Tomasiewicz, J. Valantinas, P. Van Damme, D. van de Vijver, A.J. van der Meer, D. van Santen, H. Van Vlierberghe, D. Vandijck, S. Vella, J. Videčnik-Zorman, W. Vogel, N. Weis, A. Hatzakis, Hepatitis C virus prevalence and level of intervention required to achieve the WHO targets for elimination in the European Union by 2030: a modelling study, *The Lancet Gastroenterology & Hepatology*. 2 (2017) 325–336. [https://doi.org/10.1016/S2468-1253\(17\)30045-6](https://doi.org/10.1016/S2468-1253(17)30045-6).
- [16] A. Hatzakis, V. Chulanov, A.C. Gadano, C. Bergin, Z. Ben-Ari, J. Mossong, I. Schréter, O. Baatarkhuu, S. Acharya, I. Aho, A.C. Anand, M.I. Andersson, V. Arendt, P. Arkkila, K. Barclay, F. Bessone, S. Blach, N. Blokhina, C.R. Brunton, G. Choudhuri, L. Cisneros, E.A. Croes, Y.A. Dahgwahdorj, O. Dalgard, J.R. Daruich, N.R. Dashdorj, D. Davaadorj, R.J. de Knegt, M. de Vree, C. Estes, R. Flisiak, E. Gane, E. Gower, W. Halota, C. Henderson, P. Hoffmann, J. Hornell, D. Houlihan, S. Hrusovsky, P. Jarčuška, D. Kershenobich, K. Kostrzewska, P. Kristian, M. Leshno, Y. Lurie, A. Mahomed, N. Mamonova, N. Mendez-Sanchez, S. Norris, E. Nurmukhametova, P. Nymadawa, M. Oltman, J. Oyunbileg, T. Oyunsuren, G. Papatheodoridis, N.

- Pimenov, N. Prabdial-Sing, M. Prins, S. Radke, A. Rakhmanova, K. Razavi-Shearer, H.W. Reesink, E. Ridruejo, R. Safadi, O. Sagalova, J.F. Sanchez Avila, R. Sanduijav, V. Saraswat, C. Seguin-Devaux, S.R. Shah, I. Shestakova, A. Shevaldin, O. Shibolet, M.O. Silva, S. Sokolov, M. Sonderup, K. Souliotis, C.W. Spearman, T. Staub, C. Stedman, E.A. Strebkova, D. Struck, V. Sypsa, K. Tomasiwicz, L. Undram, A.J. van der Meer, D. van Santen, I. Veldhuijzen, F.G. Villamil, S. Willemse, E. Zuckerman, F.R. Zuure, P. Puri, H. Razavi, The present and future disease burden of hepatitis C virus (HCV) infections with today's treatment paradigm - volume 2, *Journal of Viral Hepatitis*. 22 Suppl 1 (2015) 26–45. <https://doi.org/10.1111/jvh.12351>.
- [17] H. Razavi, I. Waked, C. Sarrazin, R.P. Myers, R. Idilman, F. Calinas, W. Vogel, M.C. Mendes Correa, C. Hézode, P. Lázaro, U. Akarca, S. Aleman, I. Balik, T. Berg, F. Bihl, M. Bilodeau, A.J. Blasco, C.E. Brandão Mello, P. Bruggmann, M. Buti, J.L. Calleja, H. Cheinquer, P.B. Christensen, M. Clausen, H.S.M. Coelho, M.E. Cramp, G.J. Dore, W. Doss, A.S. Duberg, M.H. El-Sayed, G. Ergör, G. Esmat, K. Falconer, J. Félix, M.L.G. Ferraz, P.R. Ferreira, S. Frankova, J. García-Samaniego, J. Gerstoft, J.A. Giria, F.L. Gonçalves, E. Gower, M. Gschwantler, M. Guimarães Pessôa, S.J. Hindman, H. Hofer, P. Husa, M. Kåberg, K.D.E. Kaita, A. Kautz, S. Kaymakoglu, M. Kraiden, H. Krarup, W. Laleman, D. Lavanchy, R.T. Marinho, P. Marotta, S. Mauss, C. Moreno, K. Murphy, F. Negro, V. Nemecek, N. Örmeci, A.L.H. Øvrehus, J. Parkes, K. Pasini, K.M. Peltekian, A. Ramji, N. Reis, S.K. Roberts, W.M. Rosenberg, F. Roudot-Thoraval, S.D. Ryder, R. Sarmento-Castro, D. Semela, M. Sherman, G.E. Shiha, W. Sievert, J. Sperl, P. Stärkel, R.E. Stauber, A.J. Thompson, P. Urbanek, P. Van Damme, I. van Thiel, H. Van Vlierberghe, D. Vandijck, H. Wedemeyer, N. Weis, J. Wiegand, A. Yosry, A. Zekry, M. Cornberg, B. Müllhaupt, C. Estes, The present and future disease burden of hepatitis C virus (HCV) infection with today's treatment paradigm, *Journal of Viral Hepatitis*. 21 Suppl 1 (2014) 34–59. <https://doi.org/10.1111/jvh.12248>.
- [18] G.J. Dore, J. Ward, M. Thursz, Hepatitis C disease burden and strategies to manage the burden (Guest Editors Mark Thursz, Gregory Dore and John Ward), *Journal of Viral Hepatitis*. 21 Suppl 1 (2014) 1–4. <https://doi.org/10.1111/jvh.12253>.
- [19] H. Wedemeyer, G.J. Dore, J.W. Ward, Estimates on HCV disease burden worldwide - filling the gaps, *Journal of Viral Hepatitis*. 22 Suppl 1 (2015) 1–5. <https://doi.org/10.1111/jvh.12371>.
- [20] M.D. Mary E. Wilson, Egypt Implements Successful Nationwide Program to Eliminate Hepatitis C, *New England Journal of Medicine Journal Watch*. 2020 (2020). <https://doi.org/10.1056/nejm-jw.NA51084>.
- [21] I. Waked, G. Esmat, A. Elsharkawy, M. El-Serafy, W. Abdel-Razek, R. Ghalab, G. Elshishiney, A. Salah, S. Abdel Megid, K. Kabil, M.H. El-Sayed, H. Dabbous, Y. El Shazly, M. Abo Sliman, K. Abou Hashem, S. Abdel Gawad, N. El Nahas, A. El Sobky, S. El Sonbaty, H. El Tabakh, E. Emad, H. Gemeah, A. Hashem, M. Hassany, N. Hefnawy, A.N. Hemida, A. Khadary, K. Labib, F.

- Mahmoud, S. Mamoun, T. Marei, S. Mekky, A. Meshref, A. Othman, O. Ragab, E. Ramadan, A. Rehan, T. Saad, R. Saeed, M. Sharshar, H. Shawky, M. Shawky, W. Shehata, H. Soror, M. Taha, M. Talha, A. Tealaab, M. Zein, A. Hashish, A. Cordie, Y. Omar, E. Kamal, I. Ammar, M. AbdAlla, W. El Akel, W. Doss, H. Zaid, Screening and Treatment Program to Eliminate Hepatitis C in Egypt, *New England Journal of Medicine*. 382 (2020) 1166–1174. <https://doi.org/10.1056/NEJMSr1912628>.
- [22] Hepatitis C Virus in Egypt: Interim Report From the World’s Largest National Program - Abdel- Razek - 2019 - *Clinical Liver Disease* - Wiley Online Library, (n.d.). <https://aasldpubs.onlinelibrary.wiley.com/doi/full/10.1002/cld.868> (accessed September 9, 2020).
- [23] An educate, test and treat model towards elimination of hepatitis C infection in Egypt: Feasibility and effectiveness in 73 villages - *Journal of Hepatology*, (n.d.). [https://www.journal-of-hepatology.eu/article/S0168-8278\(19\)30672-5/fulltext](https://www.journal-of-hepatology.eu/article/S0168-8278(19)30672-5/fulltext) (accessed September 9, 2020).
- [24] WHO | Guidelines on hepatitis B and C testing, WHO. (n.d.). <http://www.who.int/hepatitis/publications/guidelines-hepatitis-c-b-testing/en/> (accessed October 29, 2019).
- [25] M.P. McHugh, A.H.B. Wu, S. Chevaliez, J.M. Pawlotsky, M. Hallin, K.E. Templeton, Multicenter Evaluation of the Cepheid Xpert Hepatitis C Virus Viral Load Assay, *Journal of Clinical Microbiology*. 55 (2017) 1550–1556. <https://doi.org/10.1128/JCM.02460-16>.