

Letters to the Editor

Creutzfeldt-Jakob Disease Agent

To the Editor:

In a letter which discussed the problems of decontaminating medical equipment after contact with Creutzfeldt-Jakob disease (CJD) agent it was stated that autoclaving instruments at 121°C and 15 psi for one hour will result in sterilization.¹ This statement is dangerously misleading. Although CJD infectivity has been inactivated by such an exposure,² scrapie agent has survived this² and even more rigorous autoclaving procedures.³ Scrapie is the model for the group of unconventional transmissible diseases to which CJD belongs, and it is better to base decontamination standards for CJD on scrapie since there are cloned strains available that differ in their thermostability.⁴ Similar strains of CJD may exist but these agents have not yet been cloned into distinct strains. Consequently the current recommendation for CJD decontamination in the United States is gravity-displacement autoclaving at 132°C⁵ for one hour. This recommendation is based upon scrapie studies in hamsters²; experimental work with scrapie in mice³ formed the basis for the current UK standard which is porous-load autoclaving at 134-138°C for 18 minutes.⁶

It was also said that a one-hour exposure to 0.5% sodium hypochlorite is effective and is useful for situations where autoclaving is impossible.¹ However, even the use of undiluted domestic bleach (5% sodium hypochlorite) is not always effective against the CJD agent.⁵ Experiments with scrapie agent have shown that sodium hypochlorite can be effective if the exposure time is prolonged or the concentration is increased considerably,⁷ but this also increases the problem of metal corrosion. Chlorine-releasing disinfectants containing sodium dichlorisocyanurate might be useful because they are less corrosive but their effect on scrapie infectivity is not known.

Although as indicated in the letter, a one-hour exposure to 1N sodium hydroxide has been reported to be effective,⁷ this requires verification because contradictory data exists.

In conclusion, it is relevant to add to this information a comment relating to the safe handling of CJD material in the histopathology laboratory. CJD and scrapie infectivity in brain is little affected by formol fixation, and it has been suggested that autoclaving such brain tissue will render it safe for subsequent handling and processing.⁹ Although the architecture of formol-fixed scrapie brain tissue remains remarkably good after porous-load autoclaving at 134-138°C,¹⁰ the prior

fixation confers considerable protection against inactivation, and there is little reduction in infectivity after autoclaving at 134° for 18 minutes.¹¹

REFERENCES

1. Farr B: Transmission of Creutzfeldt-Jakob disease (letter). *Infect Control Hosp Epidemiol* 1988; 9:144.
2. Brown P, Rohwer RG, Gajdusek DC: Newer data on the inactivation of scrapie virus or Creutzfeldt-Jakob disease virus in brain tissue. *J Infect Dis* 1986; 153:1145-1148.
3. Kimberlin RH, Walker CA, Millson GC, et al: Disinfection studies with two strains of mouse-passaged scrapie agent. *J Neurol Sci* 1983; 59:355-369.
4. Taylor DM: Decontamination of Creutzfeldt-Jakob disease agent. *Ann Neurol* 1986; 20:749.
5. Rosenberg RN, White CL, Brown P, et al: Precautions in handling tissues, fluids, and other contaminated materials from patients with documented or suspected Creutzfeldt-Jakob disease. *Ann Neurol* 1986; 19:75-77.
6. Department of Health and Social Security: Management of patients with spongiform encephalopathy (Creutzfeldt-Jakob disease [CJD]). *DHSS circular DA (84) 16*:1984.
7. Brown P, Rohwer RG, Gajdusek DC: Sodium hydroxide decontamination of Creutzfeldt-Jakob disease virus. *N Engl J Med* 1984; 310:727.
8. Prusiner SB, McKinley MP, Bolton DC, et al: Prions: Methods for assay, purification, and characterization, in Maramorosch K, Koprowski H (eds): *Methods in Virology*, vol VIII. New York, Academic Press Incorporated, 1984, pp 293-345.
9. Masters CL, Jacobsen P, Kakulas BA: Letter to the editor. *J Neuropathol Exp Neurol* 1985; 44:304-305.
10. Taylor DM, McBride PA: Auto-wed, formol-fixed scrapie mouse brain is suitable for histopathological examination, but may still be infective. *Acta Neuropathol* 1987; 74:194-196.
11. Taylor DM, McConnell I: Autoclaving does not decontaminate formal-fixed scrapie tissues. *Lancet* 1988; i:1463-1464.

David M. Taylor, PhD
AFRC & MRC Neuropathogenesis Unit
Edinburgh, Scotland
continued on page 529

HOSPITAL EPIDEMIOLOGY:

New Challenges and Controversies

Hyatt Regency Baltimore Hotel
March 10 to 12, 1989

Program Committee

Elizabeth Ann Bolyard, R N M PH.
Richard A Garibaldi, M D
Walter J. Hierholzer, M D.
William Schaffner, M.D.
Timothy R Townsend, M D
Richard P Wenzel, M D

The Society of Hospital Epidemiologists of America (SHEA) and *Infection Control and Hospital Epidemiology*, are co-sponsoring a meeting. "Hospital Epidemiology: New Challenges and Controversies." The meeting will be held at the Hyatt Regency Hotel in Baltimore on March 10-12, 1989. The program will focus on three important areas: AIDS, the expanding roles of hospital epidemiology, and new problems for infection control. A brochure describing the conference and topics for discussion is included in this issue of the Journal. There will be presentations by an expert group of nationally-known contributors to the field of hospital epidemiology. There will also be a poster session for attendees who would like to participate in this conference. We encourage you to submit an abstract describing some aspect of your work in AIDS or hospital epidemiology.

We hope that you will join us in Baltimore. If you would like to present a poster, please complete the attached abstract form. For further information regarding this program or registration material, please write to:

Hospital Epidemiology:
New Challenges and Controversies
6900 Grove Road
Thorofare, New Jersey 08086-9447

Sponsored by

INFECTION
CONTROL
HOSPITAL EPIDEMIOLOGY & **SHEA**

6900 Grove Road
Thorofare, NJ 08086

Hospital Epidemiology: New Challenges and Controversies
March 10-12, 1989 Hyatt Regency Baltimore Hotel

DEADLINE FOR ABSTRACTS IS JANUARY 31, 1989

**ABSTRACT
REPRODUCTION FORM**

TO ENSURE THAT YOUR ABSTRACT IS
CONSIDERED, PLEASE COMPLETE ALL
INFORMATION BELOW.

If, in the conduct of these studies, human subjects were exposed to risks not required by their medical needs, the author affirms that the study was approved by an appropriate committee, or, if no such committee was available and informed consent was needed, it was obtained in accordance with the principles set forth in "The Institutional Guide to DHEW Policy on Protection of Human Subjects."

**First Author's Name, Address and Work
Telephone number**

Name _____

Address _____

Telephone _____

**Co-authors' Names, Addresses, and Status
in the College, if Applicable**

Name _____

Address _____

Name _____

Address _____

Name _____

Address _____

TYPE ABSTRACT HERE
BE SURE TO SINGLE-SPACE AND STAY WITHIN BORDER
A sharp typeface will help reproduction.

Keep original for your records. Please send three copies to:

Timothy R. Townsend, MD
Hospital Epidemiologist
Brady 119
The Johns Hopkins Hospital
600 North Wolfe Street
Baltimore, MD 21205

EXAMPLE

**RISK FACTORS FOR INFECTION FOLLOWING RENAL
TRANSPLANTATION: PROSPECTIVE STUDY. *D.G. Maki,*
B. Fox and *B. Storer.* University of Wisconsin, Madison, WI.**