


*guaranteed maximum
background of 3cpm/1.2ℓ*
Radiocarbon(¹⁴C) Dating System



This system is designed to measure efficiently the value of radioactivity emitted from ¹⁴C included in a sample by means of Multi-anode anti-coincidence gas proportional counter.

- Counter:
 - entire cubic content 2.2 ℓ
 - effective cubic content 1.2 ℓ
 - Background:
 - less than 3cpm/1.2ℓ in the form of acetylene gas (dead carbon) contained in the counter at 760mmHg and 25°C
 - Efficiency:
 - approx., 80% (¹⁴C)
- * Please write for our detailed information to:

TOKYO **A** TOMIC
TOKYO GENSHI KOGYO CO.,LTD.

2-12-8, HIGASHIGOTANDA, SHINAGAWA-KU
TOKYO, JAPAN
TEL. 03-441-1176

(Foreign representatives wanted)

CONTENTS

BC	<i>Evan T Williams</i> Brooklyn College Radiocarbon Dates I	1
CSM	<i>A K Lavrukina and V A Alexeev</i> Laboratory of Cosmochemistry Radiocarbon Measurements I	12
IJ	<i>T W Linick</i> La Jolla Natural Radiocarbon Measurements VII	19
MC	<i>J L Rapaire and G Hugues</i> Monaco Radiocarbon Measurements V	49
N	<i>Fumio Yamasaki, Chikako Hamada, and Tatsuji Hamada</i> RIKEN Natural Radiocarbon Measurements IX	62
QU	<i>Claude Samsen, Louis Barrette, P LaSalle, and J Fortier</i> Quebec Radiocarbon Measurements I	96
SUA	<i>Richard Gillespie</i> Sydney University Natural Radiocarbon Measurements IV	101
Tln	<i>J M Punning, R Rajamäe, M Ehrenpreis, and I Saro</i> Tallinn Radiocarbon Dates IV	111
UM	<i>D Piegras and J J Stupp</i> University of Miami Radiocarbon Dates VIII	118
WIS	<i>M M Bender, R A Bryson, and D A Bacreis</i> University of Wisconsin Radiocarbon Dates XIV	127
Ya	<i>Y Nozaki and K K Turekian</i> Yale University Geology and Geophysics Radiocarbon Dates I	138