

(HTA) study is to compare the performances of PFBD devices with the standard venipuncture to evaluate the potential benefits of introducing PFBD devices into clinical practice.

Methods. PFBD devices use microneedles that breach the stratum corneum, significantly reducing the pain perception due to the superficial skin penetration. Decision-oriented HTA method, was applied to conduct the HTA process. It is an analytical instrument that integrates the EunetHTA CoreModel with the analytic hierarchy process, to choose the best technology solution by identifying the main evaluation criteria and defining the weights of system and performance values. Eight professionals have been involved to define the evaluation criteria and to measure the two technologies' performance. As the method requires, a literature review was conducted to define the evaluation scheme represented by a multilevel decision tree composed of evaluation areas (domains) and key performance indicators (KPI).

Results. Five evaluation domains were included in the analysis (clinical effectiveness, safety, costs, organizational aspects, and technical characteristics), described by 35 KPIs. Preliminary clinical effectiveness results showed diagnostic concordance between blood samples obtained with PFBD and venipuncture. Even if the additional costs of PFBD, these devices seem to improve the safety by reducing the biological risks for operators. Moreover, considering pediatric patients, organizational aspects would benefit by the use of PFBD in terms of ease of use, compliance of patients, and time reduction for blood collection.

Conclusions. Results showed that PFBD not only have great repercussions in terms of clinical benefits, especially for pediatric patients, but also a significant impact in terms of organizational aspects.

OP54 The Early Detection And Warning System 'SINTESIS-New Technologies': A Horizon Scanning Experience In Spain

Ana Isabel Hijas-Gómez (ahijas@outlook.es),
Setefilla Luengo-Matos, Amparo Arias-Pacheco,
Begoña Rodríguez-Ortiz-de-Salazar,
Luis Maria Sánchez-Gómez and
María del Mar Polo-de-Santos

Introduction. SINTESIS-new technologies is the early warning system for new and emerging technologies of the Agencia de Evaluación de Tecnologías Sanitarias, Instituto de Salud Carlos III. SINTESIS is part of the Action Plan for the Early Detection of New and Emerging Technologies of the Spanish Network of Health Technology Assessment Agencies (RedETS). In RedETS, four agencies are responsible for the identification of innovative technologies. These agencies have been collaborating since 2016 according to the early awareness methods contained in the EuroScan Methods Toolkit. SINTESIS focuses on secondary information sources (i.e., experts and literature). This study describes the experience of SINTESIS in identifying and filtering new technologies in recent years.

Methods. Retrospective analysis of all new and emerging technologies notified by SINTESIS to RedETS since 2018. Technologies were analyzed on a year-by-year basis for their source of information, the clinical specialties involved, and whether technologies found in the identification phase were selected for further assessment.

Results. Between 2018–2020, SINTESIS identified 69 emerging and new technologies. Most of the information came from medical press news (35%), and medical web news (22%); other sources included experts (15%), licensing news search (12%), general press (12%), and scientific websites (6%). Almost 37 technologies (54%) were selected for further analysis. Reasons for exclusion included too early identification of technologies/prototypes without enough evidence (52%), technologies already implemented (28%), overlapping technologies between agencies (17%), and not being medical technologies (3%).

Conclusions. Experience suggests that news sections of general and medical journals, websites, and expert consultation are useful sources to identify new and emerging health technologies. The main limitation is that the technologies identified are often at too early a stage of development for further assessment. SINTESIS contributes, within a national horizon scanning system with other agencies, to broaden the information sources and provide useful data on early awareness of innovative technologies. Further studies are needed to assess the impact of emerging technologies detection on healthcare delivery.

OP55 Classification System For Innovative Medicines In The Pipeline: New Or Repurposed?

Dapo Ogunbayo, Diarmuid Coughlan, Ross Fairbairn and
Sola Akinbolade (sola.akinbolade@io.nihr.ac.uk)

Introduction. While various criteria exist to define or categorize innovative medicines as new or repurposed, to our knowledge there are no standardized systems that sufficiently capture the range of pipeline products. The National Institute for Health and Care Research Innovation Observatory (NIHR IO) undertakes routine horizon scanning to support health technology assessment (HTA) in England and maintains a comprehensive Medicines Innovation Database (MInD). The aim of this project is to develop a 'technology type' (new versus repurposed) classification system for application within the MInD and to provide a high-level analysis of the emergent data.

Methods. We reviewed gray literature, regulatory websites, and drug repositories to identify existing 'technology type' classification criteria. Preliminary definitions and classifications for use on the MInD were discussed, refined, and agreed by consensus. Innovative medicines on the MInD were classified as either new or repurposed based on their regulatory approval status (Marketing Authorization) using data from the electronic medicines compendium. For repurposed medicines, further classification was undertaken using abbreviated new drug application (ANDA) data from the FDA Orange Book to identify generic medicines (patency and exclusivity status). We combined a range of semi-automated and manually derived data during this process.